



Hunter D-25A



Hunter
ColorQuest XE

Brightness

We measure TAPPI (or GE) brightness using a Hunter D-25A Colorimeter, which measures brightness in accordance with TAPPI test method T-452. This method does not take fluorescence into account.

We also have a Hunter ColorQuest XE spectrophotometer with diffuse lighting and an integrating sphere for ISO brightness measurement. It includes a UV filter so it can measure the effect of optical brighteners. This instrument will measure both shade and brightness.